

Service Repair Sheet SRS68579

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Roger Thomas
Neath Port Talbot Hospital
MEMS Department
MEMS Engineer

01639862300
00012163

25/Sep/2023
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer
Signed: _____
Date: _____



Goods Out Only

Cleaned by Viamed before
returning to customer
Signed: _____
Date: _____

Notes 20/Sep/2023 Janine Gill
20/Sep/2023 Janine Gill
Emailed customer with instructions to send in
25/Sep/2023 Robert Connor
Received 1 x DB3 s/n M0004278, with battery and patient lead.

Ready For quote

CGreen 5.10.23

Repair Complete Signed

CGreen 20.10.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36045	Microstim DB3	2510000	M0004278	N

2540010 x 1 @ £60
S/N, SRS, SRN

UPS x 1 @ £12

Service Repair Sheet 68579

Contact Name	Roger Thomas
Company/ Hospital Name	Neath Port Talbot Hospital
Department	MEMS Department
Position	MEMS Engineer
General Phone	01639862300
Opera Account	00012163
Order Number	92394145
Date Received	25/Sep/2023
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN36045	2510000	M0004278	Microstim DB3	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	2540010
Qty: 1	2520000
Qty: 1	2530004
Qty: 1	2530012

The Microstim DB3 has been repaired. The control knob, battery cover and patient lead have been replaced.
