

EU Declaration of Conformity

This declaration of conformity is issued under the sole responsibility of the manufacturer in accordance with Regulation 2017/745/EU Article 120 and Directive 93/42/EEC in its currently valid version.

Product: **Oxygen Sensor**
Basic UDI-DI: **4036616OOMCK**
Type: **OOM111-M**

Classification: **Class IIa**
(Annex VIII)
CE Mark: 

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339 München, Germany

Conformity Assessment Process: Annex II, section 3 of Directive 93/42/EEC.

Particular product standard applied: ISO 80601-2-55

Issued by: **Honeywell Healthcare Solutions GmbH**
Eudamed-SRN: DE-MF-000004924
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from: 21.11.2022
Valid until: 26.05.2024

Place, Date: Wismar, 21.11.2022

Authorized
Signature:


ppa. Marko Sprössel


ppa. Steffen Wickner