

# EC DECLARATION OF CONFORMITY

Name and address of the  
manufacturer:

**Shenzhen Viatom Technology Co., Ltd.**  
**4E,Building 3, Tingwei Industrial Park,**  
**No.6 Liufang Road, Block 67, Xin'an Street,**  
**Baoan District, 518101 Shenzhen, P.R.China**

Name and address of Authorized  
Representative:

**MedNet EC-REP GmbH**  
**Borkstrasse 10 , 48163 Muenster,Germany**  
**Telefon: +49 251 32266-61**  
**Telefax: +49 251 32266-22**

We declare under our sole responsibility that

the medical device:

**Pulse Oximeter**

**Model: O2 RING, Wear O2**

UMDNS

**17148**

of class:

**Class IIa**

according to annex IX of directive 93/42/EEC  
meeting all the provisions of the Directive  
93/42/EEC, as amended by Directive  
2007/47/EEC, which apply to it.

Conformity assessment procedure:

**MDD 93/42/EEC Annex II excluding (4)**

Registration No.:

**HD 601373560001**

Notified Body:

**TÜV Rheinland LGA Products GmbH**  
**Tillystraße 2**  
**90431 Nürnberg**  
**Deutschland**  
**CE 0197**

Shenzhen, 2021/03/22  
Place, date

General Manager Zhou Saixin  
Name and function

