



DECLARATION OF CONFORMITY

This is to confirm that the

Therapy Equipment Diamond Range Suction Unit

<u>Part No.</u>	<u>Description</u>
7701	High Direct Diamond Suction Unit with BS Probe
7705	Rail Mounted High Diamond Suction Unit
7706	V Dovetail Mounted High Diamond Suction Unit
7715	High Direct Diamond Suction Unit with 1.8L Receiver Jar
7720	Low Direct Diamond Suction Unit with BS Probe
7725	Rail Mounted Low Diamond Suction Unit
7726	V Dovetail Mounted Low Diamond Suction Unit
7735	Low Direct Diamond Suction Unit with 1.8L Receiver Jar
7740	Thoracic Direct Diamond Suction Unit with BS Probe
7745	Rail Mounted Thoracic Diamond Suction Unit

Conforms to the EC Declaration of Conformity regarding Medical Devices as outlined in Annex II MDD 93/42 EEC consolidated 2007/47/EC of the Official Journal of the European Communities.

Therapy Equipment ensure and declare that the above products which have been classified as Class IIA (Annex IX Rule 11) are manufactured in conformity with technical documentation (Annex VII, Section 3) and meet the directive which apply to them.

Therapy Equipment declare that the product conforms to ISO 10079-3:2009, and that the Quality and Production Standards for the products conform to BS EN ISO 13485:2012.

The GMDN Code of the Product is 13325 (13332 – Thoracic).

Steve Munn
General Manager
Therapy Equipment Ltd
12th March 2013