

Service Repair Sheet SRS68589

Contact Name

Lucy Gardener

Company/ Hospital Name

Newcastle Hospitals NHS

Department

Position

Direct Phone

01912336161

General Phone

01912336161

Opera Account

00003800

Email

lucy.gardener@nhs.net

Order Number

Date Received

02/Oct/2023

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

### VIAMEDclean

#### Goods In Only

Decontamination  
certificate provided  
by customer ☒

Cleaned by Viamed,  
if no declaration certificate  
from customer ☐

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

#### Goods Out Only

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Notes 29/Sep/2023 Kate Griffiths

29/Sep/2023 Kate Griffiths

Will send in V1000, s/n GE71743931 for serv and calib.

02/Oct/2023 Robert Connor

Received 1 x V1000 s/n GE71743931, with 4 x AA battery and grey silicone cover.

Ready For quote

CGNEEN 5-10-23

Repair Complete Signed

CGNEEN 27-10-23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36056	Foetal Simulator	1410000	GE71743931	N

1480000 x 1 @ £60

S/N, SRS, SRN

1430309 x 1 FOC

SRS, SRN

UPS > 1 £12

Service Repair Sheet 68589

Contact Name	Lucy Gardener
Company/ Hospital Name	Newcastle Hospitals NHS
Direct Phone	01912336161
General Phone	01912336161
Opera Account	00003800
Email	lucy.gardener@nhs.net
Order Number	RTDN400323067
Date Received	02/Oct/2023
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type
SRN36056	1410000	GE71743931	Foetal Simulator

Serviced  
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

The V1000 Foetal Heart Simulator has been serviced. The transducer interface cushion has been replaced. A calibration check has been performed and a certificate has been produced.

---

# Order

## BILL TO ADDRESS

ACCOUNTS PAYABLE LEVEL 2 REGENT POINT REGENT FARM ROAD  
NE3 3HD NEWCASTLE UPON TYNE,  
GLN: 5055208700004 (GLN, Endpoint ID)

Jackie Steele  
jacqueline.steele1@nhs.net

Order number

RTDN400323067

Date

2023-10-12

Validity

## Seller

### Viamed Ltd

15 Station Road  
BD207DT Cross Hills, Keighley, West Yorkshire  
United Kingdom  
GLN: 5051826000007 (GLN, Endpoint ID)  
GLN: 5051826000007 (0088)

## Delivery

### 1 910G76 MEDICAL ELECTRONICS DEPT

910G76 MEDICAL ELECTRONICS DEPT ROYAL VICTORIA INFIRMARY, QUEEN VICTORIA ROAD  
NE1 4LP NEWCASTLE UPON TYNE,  
United Kingdom

## Delivery place ID:

5055208701469

## Delivery place:

1 910G76 MEDICAL ELECTRONICS DEPT

## Receiver

### The Newcastle upon Tyne Hospitals NHS Foundation Trust

Regent Point Regent Farm Road Gosforth  
NE7 7DN Newcastle upon Tyne,  
United Kingdom  
GLN: 5055208700004 (GLN, Endpoint ID)  
GLN: 5055208700004 (0088)

Line	Product	Description
------	---------	-------------

#1	EME	SERVICING, CALIBRATION AND DELIVERY OF FOETAL HEART SIMULATOR V1000 SERIAL NUMBER GE
----	-----	--