

Service Repair Sheet SRS68576

Contact Name

Katarina Hisinger

Company/ Hospital Name

Medival s.r.l

Department

Position

Direct Phone

+39 04977 5477

General Phone

Opera Account

00006557

Email

KHisinger@medival.it

Order Number

Date Received

24/Oct/2023

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RC

Date: 24/10/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 15/Sep/2023 Ryan Swaine

15/Sep/2023 Ryan Swaine

Handi+ (0111250). S/N HC86599002. Supplied on Invoice RVM139402-1 03/10/22. End User claims it can not be switched on.

24/Oct/2023 Robert Connor

Received 2 x Maxtec Handi+ s/n HC86599002, no accessories, rattles, and JA81099032, with manual, tubing, t-adapter, oxygen stem, flow diverter, and lanyard.

Ready For quote

Repair Complete Signed

C. Connor 3.11.23

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36107	Oxygen Analyser	0111250	HC86599002	Yes
SRN36108	Oxygen Analyser	0111250	JA81099032	Yes

0140050 x 1 FOC
old S/N, SRS, SRN

- SRN 36107

0111250 x 1 FOC
New S/N, SRS, SRN

0140050 x 1 FOC
S/N, SRN, SRS
NO fault found

← SRN 36108

UPS x 1 FOC or exw if they want to
send back with next order

Medival S.r.l.

35129 Padova (Italy) Via San Crispino, 33 – Magazzino/Warehouse: Via S. Crispino, 35
Tel./Phone +39 049.77.54.77 r.a. - Fax +39 049.77.58.84
e-mail: info@medival.it - www.medival.it - posta certificata: amministrazione@pec.medival.it

To:				
VIAMED LTD				
15 Station Road, Cross Hills				
Keighley, West Yorkshire, BD20 7DT, UK				
PROFORMA INVOICE				
16/10/2023			N. 03	
FOR THE ATTN. OF RYAN SWAINE				
REF: RETURN OF GOODS				
ITEM	DESCRIPTION	Q.TY	UNIT PRICE	LINE TOTAL PRICE
990111250	ANALIZZATORE DI OSSIGENO HANDI+	2	€ 10,00	€ 20,00
RETURN OF GOODS				
GROSS WEIGHT: 1 KG				
VALUE FOR CUSTOMS PURPOSES ONLY				
COUNTRY OF ORIGIN: UK				
SHIP BY TNT			INVOICE TOTAL AMOUNT	
			€ 20,00	
SIGNATURE				

Service Repair Sheet 68576

Contact Name

Katarina Hisinger

Company/ Hospital Name

Medival s.r.l

Direct Phone

+39 04977 5477

Opera Account

00006557

Email

KHisinger@medival.it

Date Received

24/Oct/2023

Booked in By

Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN36107	0111250	HC86599002	Oxygen Analyser	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0111250

The Maxtec Handi+ Oxygen analyser has been replaced under warranty with SN:HC86599029.

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN36108	0111250	JA81099032	Oxygen Analyser	

No Fault Found

Time :0 Hour(s)

The Maxtec Handi+ analyser has been tested. The unit powered up as expected. The unit calibrated in air and Oxygen. No fault was found.
