

Service Repair Sheet SRS68597

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Vaisakh Sasikumar

Royal Brompton and Harefield

Clinical Engineering Department

Technical Support Technologist

01895823737

00002790

V.Sasikumar@rbht.nhs.uk

16/Oct/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only


Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____



Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 09/Oct/2023 Kate Griffiths

09/Oct/2023 Kate Griffiths

Faulty MySign O, 0111276, s/n 101312 - faulty display, so when trying to turn on the alarms can be heard but no display, battery has been replaced.

10/Oct/2023 Steve Hardaker

Additional info: Manufactured Jan 16. Envitec suggested it could be that the battery it's definitely not as he bought a new one from us and replaced it, did a full charge and it's doing the same. Also cant be the keypad as it turns on and off and he can hear the alarms, looks like it's a display fault. SN said we can quote for repair without any evaluation fee as he wants to get more info on this failure.

16/Oct/2023 Robert Connor

Received 1 x MySign O s/n 101312, with coiled cable, OOM111 s/n A122032, flow diverter, and blue tee.

Ready For quote

CANEN 6.12.23

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36068	Oxygen Monitor	0111276	101312	22
SRN36069	Oxygen Sensor	0111560	A122032	

0140000 x 1 1/4 hr

S/N, SRS, SRN

UPS x 1 @ £12.

Royal Brompton Hospital

Sydney Street
London SW3 6NP

Main switchboard: +44 (0)20 7352 8121

DIRECT TEL: 0207 351 4467

CLINICAL ENGINEERING DEPARTMENT

11th October 2023

Viamed Ltd
Repairs Team SRS68597
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT

Dear Sirs

Please find enclosed item which are due for Repair.

MySignO Oxygen Monitor - SN: 101312

I have also enclosed the Declaration of Contamination Status.

Please can you kindly email me costs on s.fontaine@rbht.nhs.uk.

Kindly please return to myself at the address below:-

Shalarie Fontaine
Clinical Engineering Department
Royal Brompton Hospital
Sydney Street
London SW3 6NP

Many thanks



Shalarie Fontaine
Administrator
s.fontaine@rbht.nhs.uk

ENC