

All sensors passed QA tests in ambient air, 100% oxygen, linearity error tests, as well as zero offset voltage and T90 response times. All sensors meet the published specifications.
08/Dec/2023 Steve Nixon
Address for Stephen Blatherwick: National Hyperbaric Centre, 123 Ashgrove Road West, Aberdeen, AB16 5FA.

| | |
|--------|-------------------|
| 105243 | Return Unrepaired |
| 105244 | Credited |