

Service Repair Sheet SRS68622

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Manuela Kellner

Medizintechnik

Purchase Dept

+43/1/292 66 42 29

+49 38416290

00007579

Manuela.Kellner@habel-medizinte

20/Nov/2023

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☐

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RS

Date: 20/11/23

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 06/Nov/2023 Janine Gill

06/Nov/2023 Janine Gill

The customer emailed for an SRS Number for a V1000 to be sent back for an annual calibration.

SRS sent via email, with instructions on how to send it to us.

Serial number - PR02315A11.

20/Nov/2023 Robert Connor

Received 1 x V1000 s/n PR02315A11, with 4 x AA battery and blue fabric carry case. Unit has been cleaned with Clinell wipes at Goods In but is still dirty with dried-on ultrasound gel, unit needs to be cleaned before repair.

Customer needs to be reminded that units should be cleaned before sending in and must have a certificate of decontamination status.

Ready For quote

CGreen 22.11.23

Repair Complete Signed

CGreen 15.12.23

| SRN | Equipment | Stock Ref | Serial Number | Warranty |
|----------|------------------|-----------|---------------|----------|
| SRN36142 | Foetal Simulator | 1410000 | PR02315A11 | N |

1480000 x 1 €56.70
S/N, SRS, SRN

1430309 x 1 FOC
SRS, SRN

Ins ?

UPS x 1

32x24x12cm 1kg.

VIAMED

RETURN acc. to SRS68622

15 Station Road, Cross Hills, Keighley
BD20 7DT WEST YORKSHIRE
ENGLANDHellmut Habel Gesellschaft m.b.H.
Ignaz-Koeck-Strasse 20
1210 VIENNA
AUSTRIA / EUROPE
Tel.: + 43 / 1 / 292 66 42 - 0
Fax: + 43 / 1 / 290 12 45
einkauf@habel-medizintechnik.at**Enquiry no. L083358**

Date: 06.11.2023

Your phone no. +44 1535 634542
Your fax no. +44 1535 635582
Your supplier no. 301782**DELIVERY DATE**Contact: Manuela Kellner
Our Customer No.: 10007025
Page: 1 / 1

As agreed we return for annual calibration acc. to SRS68622

Please let us have your confirmation. Thank you.

| Part No. | Description | Qty. | Unit | Price | % | Total |
|-----------|--|------|------|-------|---|-------|
| 99RETOURE | Viamed 1410000; HEART SIMULATOR V100; SN 1,0 PR02315A11 (99RETOURE our ref.) *** anual calibration acc. to SRS68622 | | | 40,38 | | 40,38 |

"Steuerfreie innergemeinschaftliche Lieferung
gem. Art.6, Abs.1, iVm Art 7 UStG"

*** Proforma Rechnung: L080440***

Price for custom clearance: 485€

| Order Value | Others | | Base | Tax % | Tax Value | Total |
|-----------------|--------|--|-------|-------|-----------|-------|
| 40,38 | 0,00 | | 40,38 | 0,00 | | 40,38 |
| Payment: net 30 | | | | | EUR | 40,38 |

www.habel-medizintechnik.at

Service Repair Sheet 68622

| | |
|------------------------|---|
| Contact Name | Manuela Kellner |
| Company/ Hospital Name | Medizintechnik |
| Department | Purchase Dept |
| Direct Phone | +43/1/292 66 42 29 |
| General Phone | +49 38416290 |
| Opera Account | 00007579 |
| Email | Manuela.Kellner@habel-medizintechnik.at |
| Order Number | No PO |
| Date Received | 20/Nov/2023 |
| Booked in By | Robert Connor |

| | | | | |
|----------------|---------|------------|------------------|--|
| Repair | Ref | S/N | Equipment Type | |
| SRN36142 | 1410000 | PR02315A11 | Foetal Simulator | Complete - Repaired Time :0 Hour(s) |
| Parts Replaced | | | | |
| | Qty: 1 | 1480000 | | |
| | Qty: 1 | 1430309 | | |

V1000 has had transducer interface cushion replaced and calibration has been checked
