

Service Repair Sheet SRS68667

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Adam Kevan

Whiston Hospital

Medical Engineering

Medical Engineer

01512904562

01514261600

00002152

Adam.Kevan@sthk.nhs.uk

09/Jan/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

☐

Cleaned by Viamed,
if no declaration certificate
from customer

☐

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 04/Jan/2024 Steve Hardaker

04/Jan/2024 Steve Hardaker

Returning a pair of R22-Vi cells that are failing calibration in an incubator.

Serial numbers V106693 and V106692.

Invoice Date: 18 May 2023. Still in warranty.

09/Jan/2024 Robert Connor

Received 2 x R22Vi s/n V106692, V106693.

Ready For quote

Repair Complete Signed

CAneen 10-1-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36211	Oxygen Sensor	0110072	V106692/V106693	Yes

0140013 x 1 FOC
old SIN, SRS, SRN

0110072 x 1 FOC
new SIN, SRS, SRN

UPS x 1 FOC

Service Repair Sheet 68667

Contact Name	Adam Kevan
Company/ Hospital Name	Whiston Hospital
Department	Medical Engineering
Position	Medical Engineer
Direct Phone	01512904562
General Phone	01514261600
Opera Account	00002152
Email	Adam.Kevan@sthk.nhs.uk
Date Received	09/Jan/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN36211	0110072	V106692/V106693	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1 0110072

Replaced under warranty with S/N V107438 / V107439
