

Fwd: O2 Analyser Repair

1 message

Main Account <office@viamed.co.uk>
To: Kate Griffiths <kate.griffiths@vmsecure.me.uk>

23 January 2024 at 11:12

----- Forwarded message -----
From: **TJAGUNS, Paul (COUNTY DURHAM AND DARLINGTON NHS FOUNDATION TRUST)**
<paul.tjaguns@nhs.net>
Date: Tue, 23 Jan 2024 at 11:10
Subject: O2 Analyser Repair
To: orders@viamed.co.uk <orders@viamed.co.uk>
Cc: WATSON, David (COUNTY DURHAM AND DARLINGTON NHS FOUNDATION TRUST)
<david.watson28@nhs.net>, TJAGUNS, Paul (COUNTY DURHAM AND DARLINGTON NHS FOUNDATION TRUST)
<paul.tjaguns@nhs.net>

Hi Kate,

We spoke yesterday about O2 analyser repair. Please could I get quote / return number for the potential repair of the following;

Model	Trust Asset Number	Device Serial Number	Reported Fault
MX-300	243208	255633	Does not recognise sensor or cable, alternatives have been tried. Display shows 00.0

Separately, please could I obtain quotes / return numbers for the potential repair of the following;

Model	Trust Asset Number	Device Serial Number	Reported Fault
MX-300	283388	None visible	Does not recognise sensor or cable, alternatives have been tried. Display shows 00.0
MX-300	No asset found	338493	Does not recognise sensor or cable, alternatives have been tried. Display shows 00.0
MX-300	241061	None visible	Does not recognise sensor or cable, alternatives have been tried. Display shows 00.0

MX-300	287649	338494	Does not recognise sensor or cable, alternatives have been tried. Display shows 00.0
--------	--------	--------	--

Regards,

Paul.

CDD Services - Estates & Facilities

Paul Tjaguns

Clinical Engineering Technician

Tel: 01325 743048

Internal Ext: 43048

paul.tjaguns@nhs.net

Synchronicity
Care Limited

Estates | Facilities | Procurement

A wholly owned subsidiary of
County Durham and Darlington NHS Foundation Trust

Clinical Engineering Department

Darlington Memorial Hospital

Hollyhurst Road
Darlington
DL3 6HX

This message may contain confidential information. If you are not the intended recipient please:

i) inform the sender that you have received the message in error before deleting it; and

ii) do not disclose, copy or distribute information in this e-mail or take any action in relation to its content (to do so is strictly prohibited and may be unlawful).

Thank you for your co-operation.

NHSm ail is the secure email, collaboration and directory service available for all NHS staff in England. NHSm ail is approved for exchanging patient data and other sensitive information with NHSm ail and other accredited email services.

For more information and to find out how you can switch visit [Joining NHSm ail – NHSm ail Support](#)

https://mail.google.com/mail/u/1/?ik=b5871a5be0&view=pt&search=all&permthid=thread-f:1788879445819941740%7Cmsg-f:178887944581994...

2/2