



R22ATEC Oxygen Sensors

Tested in 16% CO₂/N₂ gas

By Michael Gonzalez

Oct 04, 2004

INTRODUCTION:

The objective of this report is to determine if exposure to short and long durations affects the offset requirement of the R22ATEC O₂ sensor. The offset specification requirement is $\leq 0.1\%$ O₂ in ≤ 30 seconds for a step change from air to 0% O₂.

TESTS:

Five (5) R22ATEC O₂ sensors (TAI PN C44611-R22ATEC) were taken out of stock for the 2 tests as described below. First, all sensors were connected in series using nylon flow adaptors with compressed air flowing at 1 scfh flow rate, with a dwell time of 2 hours. After sensors had stabilized at room temperature, approx 24 deg C, the 5 channels on the chart recorders were calibrated. Here, 100% Full Scale (FS) on the recorder was calibrated to equal 20.9% oxygen and the speed on the recorder set at 1 cm/hour.

In test #1, the long exposure time, sensors were exposed to 16% CO₂/Nitrogen at 1 scfh flow for approx. 5-8 hours during the day. First, the speed on the recorder was changed to a faster speed at 2 cm/minute where each "cm" division equaled 30 seconds. At the same time the range on the recorder was changed from 20.9% to 2.09% oxygen where 5% full scale = 0.1% oxygen. By adjusting these 2 parameters, the offset readings at 30 seconds could then be easily measured. After the first 4 minutes, the speed was changed to 1 cm/hour. For over-night run, sensors were returned to air. This process was repeated each day between the periods of 9/22/04 to 10/4/04. Air readings and offset readings at 30 seconds were taken and documented in Lab Notebook # 237. Offset data can be seen in table #1 and air output data can be seen in table #3.

In test #2, the short exposure time, sensors were allowed to stabilize in air and then were exposed to 16% CO₂/Nitrogen at 1 scfh flow for approx. 4-9 minutes. Again, the speed on the recorder was changed to a faster speed at 2 cm/minute where each division equaled 30 seconds. At the same time the range on the recorder was changed from 20.9% to 2.09% oxygen where 5% full scale = 0.1% oxygen. Adjusting these 2 parameters, the offset readings at 30 seconds could then be easily measured. After the exposure to CO₂ gas, the sample gas was switched back to air. This process was repeated 12 times during the test day 10/4/04. Again, air readings and offset readings at 30 seconds were taken and documented in Lab Notebook # 237. Offset data can be seen in table #2 and air output data in table #3.

EQUIPMENT:

- Air, 20.9% O₂
- 16% CO₂ balance N₂, Certified Mixture No. 48-91939600-001
- 100% Nitrogen
- Nylon flow adaptors (5)
- Chart recorder (2), Linseis Models L6514B
- Digital multimeter, Fluke Model 189, Cal No. 287901-1



RESULTS:

Table 1: Test #1 (long-term exposure to 16% CO2 Exposure after being in air)

Date	Time (Hrs) exposed to Air (20.9%)	% O2 Offset after 30 sec exposure to 16% CO2					Hrs left in 16% CO2
		367138	367139	367140	367141	367142	
9/22/04	2 test started	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7
9/23/04	15	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	8-3/4
9/24/04	17	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7
9/27/04	17	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7
9/28/04	16-1/2	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7-1/2
9/29/04	16-1/2	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7-1/2
9/30/04*	17	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7
10/1/04*	18-1/2	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	5-1/2
10/4/04	64-1/2 week-end run	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	2

* Note: These 2 runs for Sn 367138, -139, -140 can be seen in sample Chart #1.

Table 2: Test #2 (short-term exposure to 16% CO2 Exposure after being in air)

Cycle No.	Time (mins) exposed to Air (20.9%)	% O2 Offset after 30 sec exposure to 16% CO2					Mins left in 16% CO2
		367138	367139	367140	367141	367142	
1	4	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4-1/2
2	7	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4-1/2
3	7	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4
4	7	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4
5	9	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4
6	7	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	9
7	8	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4-1/2
8	7-1/2	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4
9	7	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	4
10	10	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	5-1/2
11 *	8	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	7-1/2
12 *	14 longer final cycle	≤0.1	≤0.1	≤0.1	≤0.1	≤0.1	10

* Note: These 2 cycles for Sn 367138, -139, -140 can be seen in sample Chart #2.



Table 3: Daily Air Output Readings, in mV

Note: Output Specification is 7.0 – 13.0 mV.

Date	Sn 367138	Sn 367139	Sn 367140	Sn 367141	Sn 367142
9/22/04	10.149	10.828	10.965	10.709	11.136
9/23/04	10.085	10.787	10.902	10.662	11.082
9/24/04	10.067	10.748	10.861	10.606	11.053
9/27/04	10.158	10.869	11.003	10.747	11.188
9/28/04	10.154	10.853	10.987	10.705	11.160
9/29/04	10.112	10.809	10.934	10.640	11.119
9/30/04	10.032	10.748	10.868	10.572	11.069
10/1/04	10.042	10.762	10.880	10.579	11.089
10/4/04	10.042	10.754	10.880	10.556	11.102

CONCLUSION:

The results from both tests performed on the 5 R22ATEC oxygen sensors show that short-term and long-term exposure to 16% CO₂/N₂ gas did not affect the offset readings of the R22ATEC sensors. The required specification of a 0.1% O₂ reading after 30 seconds from air to zero gas (in this case, 16% CO₂/N₂ gas) was achieved each time. Chart #1 and 2 of the last 2 pages show representative short and long-term exposure traces of 3 sensors, Sn 367138, 367139 and 367140. Also, as can be seen from the data in table #3, the output readings in air (taken just before exposure to CO₂ gas) were very stable with very low output drift. Outputs in air were well within the required range of 7.0-13.0 mV.



**TELEDYNE
INSTRUMENTS**
Analytical Instruments
A Teledyne Technologies Company

Chart 1: Sample run of long-term CO₂ exposure of Sn 367138, -139 & -140.




Chart 2: Sample cycle #11 & 12 of short-term CO₂ exposure - Sn 367138, -139 & -140.

