

Wardray Premise Limited Unit 2, Mulgrave Chambers 26-28 Mulgrave Road Sutton, SM2 6LE

Tel: +44(0)20 8398 991

E-mail: sales@wardray-premise.co

Website: www.wardray-premise.co

## **DELIVERY ADDRESS**

VIAMED LIMITED

SRS68681

15 STATION ROAD

CROSSHILLS

KEIGHLEY

## DELIVERY ADVICE NO

KEIGHLEY BD20 7DT			DELIVE	RY ADVICE NO
OUR REF No.	WP51637 - RETURNS		YOUR REF No.	SRS68681
	ROYAL MAIL POST		DISPATCH DATE	26.01.24
			DESCRIPT	TON
QUANTITY	REFERENCE		DESCRIPT	ION
2	OOM111	FAULTY ENVITEC OXYGEN SENSORS		
		S/N A124567 & A124575		
		THESE SENSORS FAILED ON INSTALLATION		
		THE OXYGEN LEVELS WERE NOT ACCURATE		
		PLEASE ARRANGE REPLACEMENTS/CREDIT		
		IF YOU SHOULD HAVE ANY QUERIES PLEASE CONTACT:-		
		SARAH YEOMANS		
		EMAIL: SYEOMANS@WARDRAY-PREMISE.COM		