

Subject : AW: VM2500 Firmware v8.8
Date : Fri, 15 Aug 2014 14:18:00 +0100
Linked to : Heidi Froehlich
From : "Froehlich, Heidi" <h.froehlich@bluepoint-medical.com> (By way of info@viamed.co.uk)
To : JONATHAN (Jonathan Connor) <GoldMine User>
Cc : SNIXON (Steve Nixon) <GoldMine User>

Dear Jonathan,

Please find attached the medical VM-2500 Firmware v8.8.

In general some wording bugs were removed from the Japanese version.

However because it is one general software which generated OEM versions with switch during compilation the entire version number in increased to have less errors in software version control.

Apart from this the following changes are actually present in the VM-2500 software v8.8 compared to v8.5:

- Removed "CALIBRATION" from the Service menu. Masimo urged us to do so. Too many mistakes occur during calibration. Non of their other customers have this feature implemented in the host system (monitor)
- Removed "RESET work time" in the SERVICE menu. German health insurance complaint because they pay by "work time" of the device if leased. Now only way to reset is to update software
- Solved bug that some devices do not switch on when charging the power supply (AA, Li-Ion, adaptor) and thereafter connecting the USB cable (Connected to tolerances in PCB components)

Also find attached the veterinary software v1.1.5. The only change is:

- Solved bug that some devices do not switch on when charging the power supply (AA, Li-Ion, adaptor) and thereafter connecting the USB cable (Connected to tolerances in PCB components)

Let me know if you have questions.

Best regards,

Heidi

Von: Jonathan Connor [mailto:jonathan.connor@viamed.co.uk]
Gesendet: Freitag, 15. August 2014 14:24
An: Froehlich, Heidi
Betreff: VM2500 Firmware v8.8

Dear Heidi

Please can you send me a copy of v8.8 firmware and also the changes from 8.5

Kind regards

Jonathan Connor

<http://www.viamed.co.uk>

Email jonathan.connor@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

This message has been scanned for malware.