

## Calibration Certificate

The VN202 Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

|                                    |            |
|------------------------------------|------------|
| Calibrated voltage source          | Microcal   |
| ISO calibration index number       | CE076      |
| Serial number of the Microcal      | 6939K6     |
| Certificate number of the Microcal | BTM12439   |
| Serial number of the VN202         | 03500      |
| Certificate number of the VN202    | 1707752295 |

| Voltage input: | Target reading: | VN202 reading: |
|----------------|-----------------|----------------|
| 4.78mV         | 10%             | 9.9%           |
| 9.99mV         | 20.9%           | 20.8%          |
| 10.0mV         | 21%             | 20.9%          |
| 11.95mV        | 25%             | 24.9%          |
| 23.9mV         | 50%             | 49.9%          |
| 35.85mV        | 75%             | 74.9%          |
| 47.8mV         | 100%            | 100%           |

Tested by: M.Green

Date: 12/02/2024

Signed: