

Service Repair Sheet 68676
Contact Name Dave Bell
Company/ Hospital Name North Tyneside General Hospital
Direct Phone 0344 811 8111
Opera Account 00003859
Email david.bell2@NHCT.NHS.UK
Order Number RTFE400349668
Date Received 23/Jan/2024
Booked in By

Repair	Ref	S/N	Equipment Type	
SRN36223	1410000	PR03124A10	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.
