

Service Repair Sheet 68699

Contact Name	Artwell Bere
Company/ Hospital Name	Blackpool Victoria Hospital
Department	Medical Engineering
Position	Medical Engineer
Direct Phone	01253957799
General Phone	01253 300000
Opera Account	00000570
Email	artwell.bere@bfwml.co.uk
Date Received	15/Feb/2024
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN36271	1410000	PR02676A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---