

Equipment for Repair

Date: 12/2/24
To: VIAH

Repair Reference - 35846
Net Ref - 32568691

Vet Direct
Units 9 & 10, Gateway West
Newburn Riverside
Newcastle upon Tyne
NE15 8NZ

Tel: 0191 2642705
Email: repairs@vet-direct.com

The enclosed product has been sent to you for inspection. Please provide estimate to repair – please send estimate to repairs@vet-direct.com

Product: ENVITEC OXIGUANT S

Serial No: A111129

Accessories: —

Quantity: 1

Condition: USED

Work required: CUSTOMER SAYING WON'T GO BELOW 80
OR ABOVE 10, ENGINEER FOUND DOWN
BUTTON NOT WORKING WELL.

Warranty: ☒ Yes / No 6/10/23 PO Number: 130568

Quotation required? Yes / No

Vet Direct

...on the pulse