

CERTIFICATE OF CALIBRATION

Issued by: **Vandagraph Ltd.**

Date issued: 25 September 2014

Certificate No.: 1409240

Signed: _____

Client Sumner Manufacturing Co., Inc
Date of calibration. 25 September 2014

Instrument Sumner Argo-Naught
Serial number: 0428 - 0467

Performance Status **PASS**

The equipment detailed herein has been calibrated and now conforms to the manufacturer's published specifications. The Q.A. of the above is in strict conformity to BS EN ISO 9001-2000 . All measurements made are traceable to NAMAS.

Equipment used to carry out calibration	
Fluke Graphical Multimeter, serial number 6911002, CE080	N/A
Pico Technology Virtual Instrument ADC-200, serial number 3468, CE020	N/A
Microcal Serial Number	N/A
Oxygen BOC Medical 99.5% License 07351	Y

Setting	Oxygen Reading	Helium Reading
100% Oxygen	100	1.8
Dry Air	20.9	0.3