

Service Repair Sheet 68683

Contact Name

Christopher Courage

Company/ Hospital Name

Royal Cornwall Hospital

Department

Department of Clinical Technology

Position

Entry-Level Technician

Direct Phone

01872252488

Opera Account

00005140

Email

christopher.courage@nhs.net

Order Number

23025819

Date Received

01/Feb/2024

Booked in By

Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN36237	1410000	GE71743920	Foetal Simulator	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	1480000		
	Qty: 1	1430309		

V1000 has had transducer interface cushion replaced and calibration has been checked.

---