

# Cathy Green

Service Repair Sheet SRS68705

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Adam Kevan

Whiston Hospital

Medical Engineering

Medical Engineer

01512904562

00002152

Adam.Kevan@sthk.nhs.uk

22/Feb/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination  
certificate provided  
by customer ☐

Cleaned by Viamed  
7 no decontamination certificate  
from customer ☐

Signed \_\_\_\_\_

Date \_\_\_\_\_

Goods Out Only

Cleaned by Viamed before  
returning to customer

Signed \_\_\_\_\_

Date \_\_\_\_\_

Notes 20/Feb/2024 Kate Griffiths

20/Feb/2024 Kate Griffiths

Will return under warranty 2 x R22v s/n V106684 and V106685 "failed their calibration in an incubator"

22/Feb/2024 Robert Connor

Received 2 x R22v s/n V106684, V106685

Ready For quote

Repair Complete Signed

CCGREEN 29.2.24

| S/N      | Equipment     | Stock Ref | Serial Number   | Warranty |
|----------|---------------|-----------|-----------------|----------|
| SRS68705 | Oxygen Sensor | 0110072   | V106684/V106685 | Yes      |

0140013 x 1 @ £0  
old S/N SES, SEN

0110072 x 1 @ £0  
new S/N, SES, SEN

UPS x 1 FOC.