

# Certificate of Calibration

Certificate number: 1709654414

Tom Thumb Infant Resuscitator.

Part Number: 0310030

Serial number: SE008

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with Tom Thumb service manual.

Equipment used: Digitron Digital Manometer P200 (CE178)  
Digitron Digital Manometer P200 (CE091)  
Flow Meter (CE130)  
Snoop Liquid

Test Equipment	Test	Specification	Reading	Pass/Fail
Snoop liquid	Check all ports / connections for leaks	No bubbling	N/A	PASS
Digital Manometer (CE178)	Adjustable Valve: @ 10 lpm	Minimum: <7.8 mbar	1.7	PASS
Digital Manometer (CE178)	Adjustable Valve: @ 10 lpm	Maximum: 42.1 ±1 mbar	42.3	PASS
Digital Manometer (CE178)	Precision Valve: @ 10 lpm	Maximum: 44.1 ±1 mbar	44.9	PASS
Digital Manometer (CE178)	Pressure Gauge Test @30cmH <sub>2</sub> O	Pressure: 30cmH <sub>2</sub> O ±1 (28.4 – 30.4mbar)	29.3	PASS

Name: Phil Crossley

Date: 5<sup>th</sup> March 2024

Signed: \_\_\_\_\_