



Calibration Verification

The UltraMaxO2 Oxygen Analyser has been tested for accuracy using medical grade Oxygen and air supplied by BOC.

Medical test gases used:

Medical Oxygen (BOC)	PL 0735/5000
Purity	99.5% (min)
Batch No.	15962234

Medical Air (BOC)	PL 0735/5002R
Purity	21.0%
Batch No.	15927686

Serial number of the UltraMaxO2	JL13711013
REF	R221P11-001

Oxygen input:	UltraMaxO2 reading:
99.5%	99.9%

Air input	UltraMaxO2 reading:
21.0%	21.3%

Air Flow Test:	Pass
Pressure Test:	Pass

Certificate Number	1709656032
--------------------	------------

Tested by: M.Green

Signed: *M.Green*