

Cathy Green

Service Repair Sheet SRS68704

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Alan McInnes

West Cumberland Hospital

Medical Engineering

01228814054

01228523444

00000970

Alan.McInnes@ncic.nhs.uk

21/Feb/2024

Robert Connor

Viamed

Quote

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	☒
Cleaned by Viamed, if no decontamination certificate from customer	☐
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 19/Feb/2024 Kate Griffiths

19/Feb/2024 Kate Griffiths

V1000 comes straight on when batteries are put in, turns off only when batteries are removed. Stuck at 120bpm. No s/n apart from one starting with GE as this was last calibrated by another company.

21/Feb/2024 Robert Connor

Received 1 x V1000 s/n GE71743940

Ready For quote

CGREEN 26.2.24

Repair Complete Signed

CGREEN 6.3.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36277	Foetal Simulator	1410000	GE71743940	N

1440000 x 0.25hr @ £90 per hr.
S/N, SRS, SRN

1480000 x 1 @ £60
S/N, SRS, SRN

1430309 x 1 @ FOC
SRS, SRN

1430295 x 1 @ £35
SRS, SRN

UPS x 1 @ £12.