

Service Repair Sheet 68704

Contact Name
Company/ Hospital Name
Department
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By

Alan McInnes
West Cumberland Hospital
Medical Engineering
01228814054
01228523444
00000970
Alan.McInnes@ncic.nhs.uk
RNNN400203164
21/Feb/2024
Robert Connor

Repair	Ref	S/N	Equipment Type
SRN36277	1410000	GE71743940	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309
Qty: 1	1430295
Qty: 0.25	1440000

V1000 has had transducer interface cushion and keypad replaced. Calibration has been checked.
