

Service Repair Sheet SRS68642
Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Karine Boucher
Morgate Biomedical

0972643496
0972643496
00007218
info@morgate.fr

06/Mar/2024
Robert Connor
Viamed
Quote

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer ☒	
Cleaned by Viamed, if no declaration certificate from customer ☒	
Signed:	<u>RC</u>
Date:	<u>6/3/24</u>
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed:	_____
Date:	_____

USD DIST

Notes 04/Dec/2023 Sophie Lines

04/Dec/2023 Sophie Lines

Karine has a customer who would like to return a V1000 for repair. I have asked for a description of the fault and also the serial number. Advised the procedure for returning etc.

04/Dec/2023 Sophie Lines

Karine provided the update : the volume set up is broken. I have asked her again for the serial number

12/Dec/2023 Sophie Lines

Karine provided an update - the customer still needs the device for a week, and then they will provide the information, serial number etc

20/Dec/2023 Janine Gill

The V1000 serial number is PR0490A33

21/Feb/2024 Kate Griffiths

Update, they now have device from their customer, they will send it in

06/Mar/2024 Robert Connor

Received 1 x V1000 s/n PR0490A33, with 4 x AA battery, volume control knob is broken.

Ready For quote

CCNeen 9.3.24

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36301	Foetal Simulator	1410000	PR0490A33	N

1480000 x 1 \$ 63.40

SIN, SRS, SRN

1430309 x 1 FOC

SRS, SRN

1430297 x 1 \$ 35.40

SRS, SRN

UPS x 1 0.80kg 32x24x12cm

Catherine Spence Required Reading 31 Issues 19 Unread 14 ISO
 Tasks | **My Requests** 14.00 °C



**Morgate
Biomedical**

CID4276 Opera 00007218 Type:Company
Distributor USD

Morgate Biomedical 23 Avenue de l'Europe, Le
Newton, Chatou, , 78400, France

Tel **0972643496** USD Distributor (0006)

Proforma Credit Limit 0.00USD

VAT FR26494683741

Account Home	Main Details	Account Memos	Sales History	Stock References	Cust./Supp. Returns	Service Visits
Distributor Agreements	Documents Files	Linked Issues	Regulatory Agreements			

Use :0006 USD Distributor

Repair SRS68642 Values in Pounds @ £1: £1.00

SRN36301	Foetal Simulator	1410000			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		1480000	1	\$63.40	\$63.40
		1430309	1	\$16.25	\$16.25
		1430297	1	\$35.40	\$35.40
				Parts Cost	\$115.05

**Page updated, TRipple check prices to price
in intrastats main search**

CHECK PRICES BEFORE SENDING QUOTE

Total HT	0,00
Total TVA	0,00
Total TTC	0,00