

Service Repair Sheet SRS68724
Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Trudy Moule
County Hospital
Clinical technology
Departmental Secretary
01785 230 305
01785257731
00004840
clin.techcounty@uhnm.nhs.uk

11/Mar/2024
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only
Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒
Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer
Signed: _____
Date: _____

Notes 11/Mar/2024 Robert Connor

11/Mar/2024 Robert Connor

Returned 1 x V1000 s/n PR0495A12, with 4 x AA battery and grey silicone cover. For service.

Ready For quote

CCNEON 12.3.24

Repair Complete Signed

CCNEON 19.3.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36309	Foetal Simulator	1410000	PR0495A12	N

1480000 x 1 @ £60
S/N, SRS, SRN

1430309 x 1 FOC
SRS, SRN

UPS x 1 £12.

PURCHASE ORDER: X024444
Please quote order number on all correspondence



University Hospitals
of North Midlands

NHS Trust

DELIVER TO: VAT Registration No : GB 654 9328 08 EORI Code: GB654932808000

SUPPLIER:

VIAMED LTD
15 STATION RD
CROSS HILLS
KEIGHLEY
WYORKS
BD20 7DT

INVOICE TO:

UNIVERSITY HOSPITALS NORTH MIDLANDS
C/O ELFS Shared Services
PO Box 4418, Unit 2
Swindon, SN4 4RW
Email: elfs.205uhnm@cloud-trade.net

CLINICAL TECHNOLOGY
ROYAL STOKE HOSPITAL
UNIVERSITY HOSPITALS OF NORTH MIDLANDS
578 NEWCASTLE ROAD
STOKE ON TRENT
STAFFORDSHIRE
ST4 6QG

Enquiries to : Diane Challenger
Email : opords.mali@uhnm.nhs.uk

Vendor Number: 101276
Date: 15/03/24
Requisition Number:

LINE NO	ITEM REF	DESCRIPTION	DELIVERY	QUANTITY	UNIT OF ISSUE	UNIT PRICE	LINE VALUE
1		QUOTATION QVM148791 S/N PRO495A12	15/03/24	60.00	EACH	1.00	60.00
2		V1000 FOETAL HEART SIMULATOR SERVICE AND CHEC	15/03/24	12.00	EACH	1.00	12.00
		UPS COURIER DELIVERY					
						VAT Excl:	72.00
						Total VAT	14.40
						Order Total	86.40



Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68724

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Trudy Moule

County Hospital

Clinical technology

Departmental Secretary

01785 230 305

01785257731

00004840

clin.techcounty@uhnm.nhs.uk

X024444 FROM ROYAL STOKE

11/Mar/2024

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN36309	1410000	PR0495A12	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.