

Invoice number: RST147666
Shipment date: 7th February 2024
JFD reference: 12400242-3

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

QTY 1 FAIL > 5nT (5.02nT)

SERIAL NUMBER: 105926

Name: *Callum Walls*

Position: *QUALITY ASSURANCE*

Signed by:

On behalf of JFD



Date: 19 MAR 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com