

Invoice number: RST147665
Shipment date: 4th February 2024
JFD reference: 12400242-2

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com


Result (PASS/FAIL): QTY 6 FAIL 7 SNT

SERIAL NUMBERS: 105864 (5.05nT)
105871 (6.23nT)
105880 (5.49nT)
105884 (5.11nT)
105891 (5.31nT)
105896 (5.23nT)

Name: CALLUM WALLS
Position: QUALITY ASSURANCE ENGINEER

Signed by:

On behalf of JFD



Date: 19 MAR 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Company registered in England, No. 03116432