

CERTIFICATE OF CALIBRATION

Issued by: Viamed Ltd.
15, Station Road,
Crosshills,
Keighley,
BD207DT

Date issued: 25 November 2014

Certificate No.: 1411241

Signed:



Client Organox Ltd

Date of calibration. 25 November 2014

Oxygen Monitor Type VN202 mkII

Serial number: 00473

Procedure I.D. VM3COP40.57

Performance Status PASS

The equipment detailed herein has been calibrated and now conforms to the manufacturer's published specifications. The Q.A. of the above is in strict conformity to BS EN ISO 9001-2008 & ISO 13485-2003. All measurements made are traceable to NAMAS.

Equipment used to carry out calibration

Fluke Multimeter, serial number 07010046459 (CE0180)

Microcal Serial Number 667108/210 (CE076)

Oxygen BOC Medical 99.5% License 07351

Sensor Type: R-17VAN

Sensor Serial number: 351351 B4

| | |
|--------------------------------------|------------------------|
| 100% Oxygen | 100.1 |
| Air 20.9 | 21.2 |
| Setting on Microcal to give Oxygen % | DVM output of Microcal |
| 20.9 | 9.9 mV |
| 30.0 | 14.2 mV |
| 40.0 | 18.9 mV |
| 50.0 | 23.7 mV |
| 60.0 | 28.4 mV |
| 70.0 | 33.1 mV |
| 80.0 | 37.9 mV |
| 90.0 | 42.6 mV |