

Declaration of conformity

Description	22S Oxygen sensor P/N 3013
Part Number	0110301
Manufacturer	Teledyne Analytical Instruments 16830 Chestnut Street City of Industry CA 91748 USA
Specification	TAI specification R-22A R-22MED or R-22S
Standards	Applicable sections of Bar-90 Test Analyser system Specifications April 1966 PTB-A 1810 or EN60601-1 3 rd edition EN60601-1-2 EN14971-2012 ENISO 80601-2-55:2011

This product is a relabelled R-22 type sensor to be used only in surface analysers

We Vandagraph Ltd declare that the above mentioned product meets the above standards

Signature	
Name	
Title	Director
Date	