

Date	Part number	Part number	Serial number	Output	Zero offset voltage
	VST	JFD		mV	µV
19-03-2024	8010050	DB200381	105935	7.7	26
19-03-2024	8010050	DB200381	105936	7.6	44
19-03-2024	8010050	DB200381	105937	7.7	57
19-03-2024	8010050	DB200381	105938	8.1	40
19-03-2024	8010050	DB200381	105939	7.9	59
19-03-2024	8010050	DB200381	105940	8.1	56
19-03-2024	8010050	DB200381	105941	8.0	53
19-03-2024	8010050	DB200381	105942	8.0	55
19-03-2024	8010050	DB200381	105943	8.0	56
19-03-2024	8010050	DB200381	105944	7.8	62
19-03-2024	8010050	DB200381	105945	7.9	62
19-03-2024	8010050	DB200381	105946	8.0	62
19-03-2024	8010050	DB200381	105947	7.9	50
19-03-2024	8010050	DB200381	105948	7.9	55
19-03-2024	8010050	DB200381	105949	8.0	58
19-03-2024	8010050	DB200381	105950	7.8	50
19-03-2024	8010050	DB200381	105951	8.0	56
19-03-2024	8010050	DB200381	105952	8.0	52
19-03-2024	8010050	DB200381	105953	7.7	53
19-03-2024	8010050	DB200381	105954	7.9	33
19-03-2024	8010050	DB200381	105955	8.0	50
19-03-2024	8010050	DB200381	105956	8.0	47
19-03-2024	8010050	DB200381	105957	7.9	50
19-03-2024	8010050	DB200381	105958	8.0	68
19-03-2024	8010050	DB200381	105959	8.0	49
19-03-2024	8010050	DB200381	105960	7.9	49
19-03-2024	8010050	DB200381	105961	7.9	44
19-03-2024	8010050	DB200381	105962	8.0	69
19-03-2024	8010050	DB200381	105963	8.0	47
19-03-2024	8010050	DB200381	105964	8.0	50
19-03-2024	8010050	DB200381	105965	7.9	51
19-03-2024	8010050	DB200381	105966	8.0	52
19-03-2024	8010050	DB200381	105967	7.9	54
19-03-2024	8010050	DB200381	105968	8.0	52
19-03-2024	8010050	DB200381	105969	8.1	50
19-03-2024	8010050	DB200381	105970	8.0	36