



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: VIAMED/QUOTATION

1 message

Shirland Trading <shirlandtrading@gmail.com>

27 March 2024 at 08:08

To: Kate Griffiths <kate.griffiths@viamed.co.uk>

Good Morning Kate,

I hope you are well.

Kindly quote us for the following items and confirm the stock availability:

011019 = 10 QTY

0110021= 43 QTY

Many thanks,

Sue

--

Shirland General Trading LTD

First Floor

[182-184 Edgware Road](#)

[W2 2DS](#)

[London](#)

2 attachments

Main OEM: Dräger
OEM Sensor Part No.: 83 620 30
Output Connector: Molex, 3 pin plug
Sensor Output: 9.0 - 13.0 mV
Sensor Life: ≥ 500,000 % oxygen hours
Response Time: ≤ 5 seconds
Warranty: 12 months

IMG-20240327-WA0001.jpg

26K

Main OEM: Viamed
Application: Incubators
Output Connector: Phone jack, 3.5 mm
Sensor Output: 9.0 - 14.0 mV
Sensor Life: ≥ 1,000,000 % oxygen hours
Response Time: ≤ 12 seconds
Warranty: 12 months

IMG-20240327-WA0000.jpg

23K