

EC Declaration of Conformity

We hereby declare under our sole responsibility
that the product

GAS SAMPLING LINES WITH O2 SUPPLY FOR USE WITH SIDESTREAM GAS EXCHANGE MONITORING DEVICES

Product No:

7090121005, 7090121006, 7090121007, 7090121011, 7090121012, 7090121013,
7090121016, 7090121017, 7090121020, 7090121021, 7090121022, 7090121023,
7090121024, 7090121025, 7090121030, 7090121031, 7090121032, 7090121036,
7090121037, 7090121038, 7090121041, 7090121042, 7090121045, 7090121046,
7090121047, 7090121048, 7090121049, 7090121050, 7090121051, 7090121077,
7090121078, 7090121079, 7090121080, 7090121081, 7090121082, 7090121083,
7090121084, 7090121085, 7090121086, 7090121087, 7090121088, 7090121089,
7090121090, 7090121091, 7090121092, 7090121093, 7090121094, 7090121095,
7090121096, 7090121097, 7090121098, 7090121099, 7090121100, 7090121101,
7090121127, 7090121128

Conforms with the essential requirements of Annex I and Annex VII of the Council
Directive 93/42/EEC as amended by 2007/47/EC concerning medical devices.

In accordance with Annex IX of the Council Directive 93/42/EEC the product has
been classified as Class IIa.

Application of the CE-marking:



Notified Body:
MEDCERT GmbH
Pilatuspool 2
20355 Hamburg
Reg. No. 0482

Validity:

Date of expiry: 31 December 2015

Issuer:

bluepoint medical GmbH & Co. KG
An der Trave 15
D-23923 Selmsdorf
Germany

Place, Date:

Selmsdorf, 06. November 2014

Legally binding signature:

.....
Bernd Lindner (Managing Director)