

DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**


JFD purchase order number: 12401010-1

Serial number/s	Manufacture – Year/Month
105935 – 105970 (qty 36)	2024/03

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name:


S Nixon

Position:

Director

Date:

19th March 2024



Invoice number: RST148353
Shipment date: 25th March 2024
JFD reference: 12401010-1

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to
VST:

info@vandagraphst.com

Result (PASS/FAIL):

1x over 5nT.

Name:

Position:

S. Channing
Quality Manager

Signed by:

On behalf of JFD

Date:

4/4/24

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μ V	
19-03-2024	8010050	DB200381	105935	7.7	26	4.17
19-03-2024	8010050	DB200381	105936	7.6	44	4.30
19-03-2024	8010050	DB200381	105937	7.7	57	3.52
19-03-2024	8010050	DB200381	105938	8.1	40	4.10
19-03-2024	8010050	DB200381	105939	7.9	59	4.15
19-03-2024	8010050	DB200381	105940	8.1	56	4.96
19-03-2024	8010050	DB200381	105941	8.0	53	3.95
19-03-2024	8010050	DB200381	105942	8.0	55	3.71
19-03-2024	8010050	DB200381	105943	8.0	56	3.75 4.18
19-03-2024	8010050	DB200381	105944	7.8	62	4.18 3.62
19-03-2024	8010050	DB200381	105945	7.9	62	4.42
19-03-2024	8010050	DB200381	105946	8.0	62	4.04
19-03-2024	8010050	DB200381	105947	7.9	50	4.503
19-03-2024	8010050	DB200381	105948	7.9	55	3.75
19-03-2024	8010050	DB200381	105949	8.0	58	4.20
19-03-2024	8010050	DB200381	105950	7.8	50	3.63
19-03-2024	8010050	DB200381	105951	8.0	56	3.54
19-03-2024	8010050	DB200381	105952	8.0	52	3.60
19-03-2024	8010050	DB200381	105953	7.7	53	3.67
19-03-2024	8010050	DB200381	105954	7.9	33	4.19
19-03-2024	8010050	DB200381	105955	8.0	50	4.87
19-03-2024	8010050	DB200381	105956	8.0	47	3.78
19-03-2024	8010050	DB200381	105957	7.9	50	5.58
19-03-2024	8010050	DB200381	105958	8.0	68	3.54
19-03-2024	8010050	DB200381	105959	8.0	49	3.64
19-03-2024	8010050	DB200381	105960	7.9	49	4.08
19-03-2024	8010050	DB200381	105961	7.9	44	3.83
19-03-2024	8010050	DB200381	105962	8.0	69	4.01
19-03-2024	8010050	DB200381	105963	8.0	47	3.74
19-03-2024	8010050	DB200381	105964	8.0	50	3.72
19-03-2024	8010050	DB200381	105965	7.9	51	3.76
19-03-2024	8010050	DB200381	105966	8.0	52	4.21
19-03-2024	8010050	DB200381	105967	7.9	54	3.72
19-03-2024	8010050	DB200381	105968	8.0	52	4.25
19-03-2024	8010050	DB200381	105969	8.1	50	3.87
19-03-2024	8010050	DB200381	105970	8.0	36	4.22