

Service Repair Sheet SRS68740

Contact Name Hussein Remtulla
 Company/ Hospital Name St James University Hospital
 Department Clinical Engineering
 Position Medical Electronics Technician
 Direct Phone 01132066882
 General Phone 01132433144
 Opera Account 00002570
 Email hussein.remtulla@nhs.net
 Order Number
 Date Received 28/Mar/2024
 Booked in By Robert Connor
 Main Company Viamed
 Type Return Quote

VIAMEDclean
Goods In Only
 Decontamination certificate provided by customer ☒

Cleaned by Viamed, if no declaration certificate from customer ☒
 Signed: _____
 Date: _____

Goods Out Only
 Cleaned by Viamed before returning to customer
 Signed: _____
 Date: _____

Notes 25/Mar/2024 Kate Griffiths

25/Mar/2024 Kate Griffiths

2 DB3 for repair, s/n: H0002421 and M0004907. Decontam. cert asked for and nature of faults

26/Mar/2024 Kate Griffiths

EDIT: s/n H0002421 - works intermittently, the arrow on the dial does not seem to be in the right place. S/N: M0004907 - broken dial so can't test if working or not. Decont cert will be included in parcel.

28/Mar/2024 Robert Connor

Received 2 x microstim DB3 s/n H0002421, M0004907, with 1 x patient lead.

Ready For quote

CGREEN 5.4.24

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36344	Microstim DB3	2510000	H0002421	22
SRN36345	Microstim DB3	2510000	M0004907	22

2540010 x 2 @ £60 each

S/N's SRS, SRN'S

UPS x 1 £12