

Service Repair Sheet 68723

Contact Name	Katherine Jones
Company/ Hospital Name	South Tyneside District Hospital
Department	Clinical Engineering Dept
Position	Clinical Engineer
Direct Phone	0191 4041107
General Phone	0191 404 1000
Opera Account	00004710
Email	Katherine.Jones41@nhs.net
Order Number	RLNP400097317
Date Received	21/Mar/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN36332	0310030	VMDF03	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced  
 Qty: 1 0380000

Tom Thumb TT480 VMDF03 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN36333	0310030	VMDF04	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced  
 Qty: 1 0380000

Tom Thumb TT480 VMDF04 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN36334	0310030	VMDF08	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced  
 Qty: 1 0380000

Tom Thumb TT480 VMDF08 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.