



Derek Lamb &lt;liquidgands@gmail.com&gt;

---

**AW: CE Certificate for Acorbat 200**

1 message

**Urfer Daniel, Nufer Medical AG** <d.urfer@nufer-medical.ch>

26 February 2015 at 13:28

To: Derek Lamb &lt;derek.lamb@viamed.co.uk&gt;

Dear Derek

The arm brackets and the roll stand have their own CE Certificate, because these are independent medical devices in Class I. According to the Medical Devices Directive 93/42/ECC , we combine different medical devices to a medical system. For the whole system no separate CE Certificate is issued.

Please find enclosed all relevant Certificates: The Certificates from are *Ondal* and *Heinen+Löwenstein* are relevant the older versions. The Certificates from *MZ Lieberec a.s.* are relevant for the current deliveries.

All of these documents are also available on the CD that comes with each Ceratherm.

I hope this is helpful for you.

Kind regards

Daniel

Daniel Urfer

General Manager

**Nufer Medical AG**

Morgenstrasse 148

CH-3018 Bern

Phone +41 (0)31 958 66 66

Fax +41 (0)31 951 46 73

[d.urfer@nufer-medical.ch](mailto:d.urfer@nufer-medical.ch)

[www.nufer-medical.ch](http://www.nufer-medical.ch)

**Von:** [liquidgands@gmail.com](mailto:liquidgands@gmail.com) [mailto:[liquidgands@gmail.com](mailto:liquidgands@gmail.com)] **Im Auftrag von** Derek Lamb

**Gesendet:** Donnerstag, 26. Februar 2015 13:53

**An:** Urfer Daniel, Nufer Medical AG

**Betreff:** CE Certificate for Acorbat 200

Dear Daniel,

Normally Steve Nixon would deal with these enquiries however he is currently away from the office.

I've been asked if the Nufer CE Certificate includes the arm brackets, or if the arm brackets have their own CE Certificate.

I only store / track the main Nufer CE Certificate and have not been asked this before.

Kind regards

Derek I Lamb

Managing Director - Viamed Ltd.

<http://www.viamed.co.uk>

Email [derek.lamb@viamed.co.uk](mailto:derek.lamb@viamed.co.uk)

Tel: [+44 \(0\)1535 634542](tel:+44(0)1535634542)

Fax: [+44 \(0\)1535 635582](tel:+44(0)1535635582)

Company registered in England, No. 01291765.

E-mail Disclaimer

-----

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call [+44 \(0\)1535 634542](tel:+441535634542)

-----

---

**5 attachments**

**16.025.B\_Konformitätserklärung\_Ceratherm\_600-3.pdf**  
463K



**CE-d-Acrobat2000\_2013.pdf**  
235K



**DoC\_H+L\_Isotherm Fahstativ (3).pdf**  
297K



**Declaration\_of\_Conformity\_EPS 712 COMb AJ\_Aug 2012.pdf**  
437K



**Declaration\_of\_Conformity\_EPS 715.pdf**  
482K