



Ambient Pressure Diving LTD Water-ma-Trout Helston, Cornwall United Kingdom TR13 OLW info@apdiving.com

Mr S Nixon VANDAGRAPH SENSOR TECHNOLOGIES 15 STATION CROSSHILLS KEIGHLEY BD207DT

27-02-2015

REF: CAR 573

Hi Steve

Please find enclosed 3 cells:-

which is from a customer and has no output. 102970 which had no output on inspection here. 101981

which was found to have a low reading here of 7.1mV 106237

Best regards

Terry Fisher Chief Engineer

> COMMERCIAL MILITARY TECHNICAL

ISO 9001