

Service Repair Sheet 68716

Contact Name	Matthew Jameson
Company/ Hospital Name	Sunderland Royal Hospital
Department	Clinical Engineering
Position	Senior Clinical Engineering Technician
Direct Phone	01915699933
Opera Account	00004992
Email	matthew.jameson@nhs.net
Order Number	RLNP400097956
Date Received	06/Mar/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN36295	0310030	VMA0086	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA0086 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN36296	0310030	VMA9001	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA9001 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN36297	0310030	VMA0105	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA0105 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN36298	0310030	VMA0106	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA0106 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
