Service Repair Sheet 68716

Contact Name

Company/ Hospital Name

Department

Position Direct Phone

Opera Account

Email

Order Number

Date Received Booked in By

Matthew Jameson

Sunderland Royal Hospital Clinical Engineering

Senior Clinical Engineering Technician

01915699933 00004992

matthew.jameson@nhs.net

RLNP400097956 06/Mar/2024 Robert Connor

Repair

Ref

S/N SRN36299 0310030 VMA0110 Equipment Type

Tom Thumb Complete - Repaired Time: 0 Hour(s)

Parts Replaced

0380000 Qty: 1

Tom Thumb TT480 VMA110 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair

Ref

S/N SRN36300 0310030 VMA0115 Equipment Type Tom Thumb

Complete - Repaired Time: 0 Hour(s)

Parts Replaced

Qty: 1 0380000

Tom Thumb TT480 VMA115 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.