

Service Repair Sheet SRS68740

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Hussein Remtulla
St James University Hospital
Clinical Engineering
Medical Electronics Technician
01132066882
01132433144
00002570
hussein.remtulla@nhs.net

28/Mar/2024
Robert Connor
Viamed
Quote

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	<input checked="" type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input checked="" type="checkbox"/>
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 25/Mar/2024 Kate Griffiths

25/Mar/2024 Kate Griffiths

2 DB3 for repair, s/n: H0002421 and M0004907. Decontam. cert asked for and nature of faults

26/Mar/2024 Kate Griffiths

EDIT: s/n H0002421 - works intermittently, the arrow on the dial does not seem to be in the right place. S/N: M0004907 - broken dial so can't test if working or not. Decont cert will be included in parcel.

28/Mar/2024 Robert Connor

Received 2 x microstim DB3 s/n H0002421, M0004907, with 1 x patient lead.

Ready For quote

CANEEN 5.4.24

Repair Complete Signed

CANEEN 19.4.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36344	Microstim DB3	2510000	H0002421	22
SRN36345	Microstim DB3	2510000	M0004907	22

2540010 x 2 @ £60 each

S/N's SES, SRN'S

UPS x 1 £12