

Viamed Limited - 15 Station Road - Cross Hills Keighley · West Yorkshire · BD20 7DT · United Kingdom Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582 Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68755

Contact Name

Company/ Hospital Name

Department Position

Direct Phone General Phone Opera Account

Email

Date Received Booked in By

Roy Gilbride

Northampton General Hospital

EBME Department

Senior Biomedical Engineer

01604545440 01604 634700 00003880

roy.gilbride@nhs.net

18/Apr/2024 Robert Connor

Repair SRN36364 Ref

S/N

1410000 PR03495A11

Equipment Type Foetal Simulator

Under Warranty Complete - Repaired Time: 0 Hour(s)

Parts Replaced

Qty: 1 1480000 Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.