



Service Repair Sheet 68756

Contact Name	Purchasing
Company/ Hospital Name	Eden Medical (UK) Ltd
Direct Phone	01314406500
General Phone	01314406500
Opera Account	00012178
Email	<a href="mailto:info@edenmedical.co.uk">info@edenmedical.co.uk</a>
Order Number	DPO-02061
Date Received	22/Apr/2024
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN36370	1410000	PR02643A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.

---