

Invoice number: RST148358
Shipment date: 5th March 2024
JFD reference: 12401010-4

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

QTY 14 FAIL 7 SNT, AS HIGHLIGHTED NERLEAF
5-01MT TRN 6-35NT

Name: CALLUM WATTS

Position: QUALITY ASSURANCE ENGINEER

Signed by:

On behalf of JFD



Date: 30 APR 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μ V	
05-04-2024	8010050	DB200381	106043	7.8	26	4.00
05-04-2024	8010050	DB200381	106044	7.8	43	5.51
05-04-2024	8010050	DB200381	106045	7.7	39	4.88
05-04-2024	8010050	DB200381	106046	7.8	31	4.43
05-04-2024	8010050	DB200381	106047	7.6	30	4.50
05-04-2024	8010050	DB200381	106048	7.7	51	5.10
05-04-2024	8010050	DB200381	106049	7.9	32	5.22
05-04-2024	8010050	DB200381	106050	7.7	29	4.64
05-04-2024	8010050	DB200381	106051	7.8	34	4.16
05-04-2024	8010050	DB200381	106052	7.9	23	3.61
05-04-2024	8010050	DB200381	106053	7.8	26	4.55
05-04-2024	8010050	DB200381	106054	7.8	31	6.35
05-04-2024	8010050	DB200381	106055	7.7	37	5.27
05-04-2024	8010050	DB200381	106056	7.7	18	4.01
05-04-2024	8010050	DB200381	106057	7.7	33	4.47
05-04-2024	8010050	DB200381	106058	7.7	33	4.44
05-04-2024	8010050	DB200381	106059	7.7	47	5.08
05-04-2024	8010050	DB200381	106060	7.7	28	3.94
05-04-2024	8010050	DB200381	106061	7.6	38	5.19
05-04-2024	8010050	DB200381	106062	7.7	16	6.04
05-04-2024	8010050	DB200381	106063	7.7	32	5.50
05-04-2024	8010050	DB200381	106064	7.7	43	4.05
05-04-2024	8010050	DB200381	106065	7.8	19	4.50
05-04-2024	8010050	DB200381	106066	7.7	42	3.77
05-04-2024	8010050	DB200381	106067	7.8	30	4.58
05-04-2024	8010050	DB200381	106068	7.7	23	4.00
05-04-2024	8010050	DB200381	106069	7.8	15	5.01
05-04-2024	8010050	DB200381	106070	7.8	32	4.64
05-04-2024	8010050	DB200381	106071	7.6	13	5.50
05-04-2024	8010050	DB200381	106072	7.8	38	5.38
05-04-2024	8010050	DB200381	106073	7.8	26	5.21
05-04-2024	8010050	DB200381	106074	7.9	33	5.64
05-04-2024	8010050	DB200381	106075	7.7	52	3.72
05-04-2024	8010050	DB200381	106076	7.7	55	4.40
05-04-2024	8010050	DB200381	106077	7.7	18	4.78
05-04-2024	8010050	DB200381	106078	7.7	27	3.61

 - < 5nT
 - 7.5nT - < 6nT
 - 7.6nT

DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**

JFD purchase order number: 12401010-4

Serial number/s	Manufacture – Year/Month
106043 – 106078 (qty 36)	2024/04

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name:


S Nixon

Position:

Director

Date:

5th April 2024



VERIFIED
CERTIFICATE

CHECKED BY **C. WALLS** DATE **30/4/24**
RECEIVER # **303632** OCR # **MULTIAZ**
WO/PO # **P01240101014**
JFD PART NUMBER **P0200381 QTY 36**
TRACEABLE MARK **RS HIGHWAY100**

22 L5nT
12 75nT - 46nT
2 76nT