
VIAMED Management Meeting - Other

Oxygen Sensor User Instructions

Issue ID #61812

Date Created 12/Aug/2015

Issued To: Derek Lamb

Completed Status: Still Outstanding

Notes:

Added by Steve Nixon sent to Derek Lamb

Statement from Envitec ref. no need for oxygen sensor user instructions. Believe we already clarified this with the auditor today.

Hi Steve,

this is the Statement from Marko:

We refer to the user manual of the device. The sensor exchange is described In this user manual. As we are the compatible product to the original, our sensor will be handled exactly the same.

Furthermore our sensors must be changed exclusively by professional user (Biomedics etc.)

I hope that could help.

Jessica
