

Service Repair Sheet 68805

Contact Name	Barry Moore
Company/ Hospital Name	Frimley Park Hospital
Department	EME Department
Direct Phone	0300 613 6332
General Phone	01344877482
Opera Account	00001740
Email	barry.moore@nhs.net
Order Number	AA018746
Date Received	05/Jun/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN36490	1410000	PR03164A10	Foetal Simulator	Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
