

Viamed Limited - 15 Station Road - Cross Hills Keighley · West Yorkshire · BD20 7DT · United Kingdom Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582 Email: Info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68796

Contact Name

Company/ Hospital Name

Department Position Direct Phone

General Phone Opera Account

Email

Order Number Date Received Booked in By

Elliot Rollings

St James University Hospital

Clinical Engineering

Medical Electronics Technician

0113 2066882 01132433144 00002570

elliot.rollings@nhs.net RR8N400478363 04/Jun/2024 Robert Connor

Repair SRN36477 2510000 H0000240

Ref

S/N

Equipment Type Microstim DB3

Complete - Repaired Time: 0 Hour(s)

Parts Replaced

Qty: 1 2540010 Qty: 1 2510000 2520000 Qtv: 1

The Microstim DB3 has been repaired and the patient lead has been replaced.