

Service Repair Sheet SRS68792

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Alex Ferguson

John Radcliffe Hospital

Clinical Engineering

01865572163

01865741166

00004040

Alex.Ferguson@ouh.nhs.uk

23/May/2024

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 22/May/2024 Kate Griffiths

22/May/2024 Kate Griffiths

DB3 not working, he has tried different leads and batteries. S/N H0000254. Catrin aware as she emailed him with service chase in April 2024

23/May/2024 Robert Connor

Received 1 x Microstim DB3 s/n H0000254, with 9v battery.

Ready For quote

CCGreen 31.5.24

Repair Complete Signed

CCNeen 26.6.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36464	Microstim DB3	2510000	H0000254	N

2540010 x 1 @ £60

S/N, SRS, SRN

UPS x 1 @ £12