

Total Net: 472.00
Total Vat: 94.40
Total: 566.40



Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68792
Contact Name
Company/ Hospital Name
Department
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By

Alex Ferguson
John Radcliffe Hospital
Clinical Engineering
01865572163
01865741166
00004040
Alex.Ferguson@ouh.nhs.uk
RTHN400305906
23/May/2024
Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN36464	2510000	H0000254	Microstim DB3	Time :0 Hour(s)

Parts Replaced

Qty: 1	2540010
Qty: 1	2510000
Qty: 1	2520000

The Microstim DB3 has been repaired and the patient lead has been replaced.